

## Converting C to F when looking at tolerance

A mistake that is common in the conversion of C to F when you look at tolerance is the fact that the conversion is not the same conversion for measured temperature.

1°C is equal to 1.8°F. Both have a different starting point, which is why when using the measured conversion formula of  $T(^{\circ}\text{F}) = T(^{\circ}\text{C}) \times 9/5 + 32$  you are adding or subtracting 32 in the process.

When looking at tolerance you should stick to the ratio of C to F or 1:1.8

For example,  $\pm 15^{\circ}\text{C}$  is equal to  $15 \times 1.8 = 27^{\circ}\text{F}$ . if we were looking at measured temperature we would be looking at.  $15^{\circ}\text{C} \times 9/5 + 32 = 59 \text{ F}$ .

HakkoUSA Knowledge Base

<https://kb.hakkousa.com/Knowledgebase/11699/Converting-C-to-F-when-looking-...>